

## TOWARDS A BETTER DIAGNOSIS AND PROGNOSIS IN PERI-IMPLANT DISEASE FROM A MICROBIOLOGICAL PERSPECTIVE. A NARRATIVE REVIEW

Giorgio Nichitean<sup>1</sup>, Ionut Luchian<sup>2\*</sup>, Ioana Martu<sup>3\*</sup>, Diana Diaconu-Popa<sup>3</sup>, Ancuta Goriuc<sup>4</sup>, Catalin Tibeica<sup>5</sup>, Oana Butnaru<sup>6</sup>, Oana Bejan<sup>7</sup>, Monica Tatarciuc<sup>3</sup>, Ramona Gabriela Ursu<sup>8</sup>

<sup>1</sup> Phd Student “Grigore T. Popa” University of Medicine and Pharmacy, Iasi-Romania.

<sup>2</sup> “Grigore T. Popa” University of Medicine and Pharmacy Iasi Romania, Faculty of Dental Medicine. Department of Periodontology.

<sup>3</sup> “Grigore T. Popa” University of Medicine and Pharmacy Iasi Romania, Faculty of Dental Medicine. Department of Implantology and Prosthetic Dentistry.

<sup>4</sup> “Grigore T. Popa” University of Medicine and Pharmacy, Iasi-Romania, Faculty of Dental Medicine, Department of Biochemistry

<sup>5</sup> “Grigore T. Popa” University of Medicine and Pharmacy, Iasi-Romania, Faculty of Dental Medicine, Department of Medical and Dental Management.

<sup>6</sup> “Grigore T. Popa” University of Medicine and Pharmacy, Iasi Romania, Faculty of Dental Medicine, Department of Biophysics.

<sup>7</sup> Postgraduate Student, “Grigore T. Popa” University of Medicine and Pharmacy, Iasi Romania, Faculty of Dental Medicine, Department of Orthodontics

<sup>8</sup> “Grigore T. Popa” University of Medicine and Pharmacy, Iasi-Romania, Faculty of Medicine, Department of Preventive Medicine and Interdisciplinarity (IX)—Microbiology

*Corresponding authors:* Ionut Luchian\*, e-mail: ionut.luchian@umfiasi.ro

Ioana Martu\*, e-mail: ioana.martu@umfiasi.ro

*#All authors have contributed equally as the first author*

**Abstract:** With the increasing popularity of dental implants among patients as a treatment alternative for edentulousness of various causes, the occurrence of implant complications is continuously rising. The need to know the latest techniques for managing these complications is vital for clinicians. There is currently no data in the literature on possible biological or mechanical complications of this approach. The aim of this review was to identify, in the specialised literature databases, established pathogenic microorganisms associated with peri-implant disease and to observe their grafting onto the implant surface after the application of implantoplasty techniques. The search was performed in databases such as PubMed, Google Academic, Web of Science and Cochrane Library and was conducted for the last thirty years (1993 - 2023). Different combinations of keywords, treatment alternatives of peri-implant pathology, as well as pathogenic microorganisms, involved in the occurrence of the disease were used for the search. We identified elements of etiopathogenicity (dental plaque and smoking), molecular biology mechanisms involved in peri-implant pathology, and we detailed possible particular developments of invasive peri-implant therapies.

Conclusions implantoplasty is a modern treatment alternative for peri-implant pathology. Using sensitive and specific methods (e.g., sequencing), numerous bacterial species associated with dental implants have been identified. In conclusion, implantoplasty requires careful prospective evaluation from microbiological point of view.

**Keywords:** *periimplantitis, periodontopathogenic microorganisms, implant surface modification, implantoplasty, periimplant disease, prognosis, diagnosis*

## 1. Introduction

Periimplantitis is an inflammatory disease characterised by increased probing depth, the presence of bleeding and/or suppuration and progressive loss of peri-implant supporting bone tissue, which may lead to loss of the dental implant [1]. It is a prevalent condition affecting a significant number of people with implants [2]. A history of periodontitis, inadequate biofilm control and lack of regular maintenance therapy are recognised as risk factors for the development of peri-implantitis [3-8]. Various therapies have been described to address peri-implantitis, including non-surgical therapies and resective or regenerative surgical therapies. In many cases, a combination of techniques is used [9].

## 2. General Aspects and Aetiopathogenesis

### 2.1 Dental plaque

Regular maintenance and adequate accessibility to achieve plaque control are essential in order to avoid recurrence of peri-implant disease with eventual loss of the implant [10-13]. Martins et al emphasise that patient maintenance and compliance were fundamental to achieving these results, as we know that compliance and regular professional control are essential to achieve peri-implant soft tissue health [10]. A recent study found that full-arch implant-prosthetic restorations have a 16.1-fold higher risk of peri-implantitis than single implant-prosthetic restorations [14] most likely due to the difficulty in performing adequate oral hygiene. [15,16]

This could indicate that fully rehabilitated patients should have an appropriate design of supra-implant prosthetic restorations and perhaps a more restrictive maintenance care programme [17].

The high success rate of dental implants has contributed to their consolidation as a therapeutic alternative for the replacement of missing teeth [18]. Despite the high success rate, inflammatory changes in peri-implant tissues due to bacterial infection are a concern for

practitioners and require attention [19]. Bacterial accumulation on the implant surface plays a central role in the etiopathogenesis of peri-implantitis, contributing to progressive loss of supporting bone and poor prognosis [20].

Therefore, the approach to peri-implantitis treatment involves eliminating implant surface contamination, attenuating the inflammatory process and halting the progression of bone loss, which will ultimately improve clinical outcomes after implantoplasty, characterised by a decrease in purulence and bleeding, as well as reduced periodontal pockets, increasing implant survival time [9].

### 2.2 Smoking as a risk factor

Smoking is an important factor that has been reported to influence treatment outcome [21,22].

However, consistent with a recent study that included 6 smokers out of a total of 25 patients, smoking had no effect on the treatment outcome of implantoplasty [23]. In the present cases, there was no indication that smoking exerted an influence on the surgical outcome after implantoplasty, as peri-implantitis resolution was observed in both smoking patients. However, the small number of smokers does not allow a firm conclusion [17].

## 3. Molecular biology mechanisms involved in peri-implant pathology

It has been shown that particulate debris resulting from specific therapeutic procedures can act as a biologically active substance and initiate an osteolytic reaction leading to failure of implant-prosthetic rehabilitations [24].

Furthermore, differences in composition, concentration, morphology and size of metal particles exert different cytotoxic effects [25]. Among these factors, particle size and concentration have a greater impact [1, 24] Most wear particles found in failed joint replacements are submicron in size (<1 μm) and are more likely to be responsible for the osteolytic response initiated and mediated by macrophages [25-35] Specifically,

macrophages can phagocytose wear debris [36-37], leading to the production of proinflammatory signalling molecules such as prostaglandin E2 (PGE2), tumour necrosis factor alpha (TNF- $\alpha$ ), interleukin-1 (IL-1), and interleukin-6 (IL-6) [38-40].

These cytokines have been shown to be stimulated when particle size falls within a certain range (0.6-4.5  $\mu$ m) [41]. To promote successful osseointegration and faster bone healing, implant manufacturers have incorporated various surface treatments, including sandblasting, acid etching, plasma spraying, anodising, etc. [42].

Higher levels of biological and immunological activity in metal particles have been observed in the size range of 1-50  $\mu$ m, and/or at the nanometre level [38-43].

However, these modifications may facilitate the release of wear particles during implant handling, placement, maintenance, peri-implant disease management, etc. This alteration of the implant surface may be influenced by fixture structure, roughness and topographic configuration [43] and could influence peri-implant progression and/or treatment [44].

#### 4. Elements of diagnosis and treatment

##### 4.1 General principles and microbiological background

Several treatment modalities and protocols have been used for the therapy of peri-implant pathology, including mechanical debridement, use of antiseptics, local or systemic antibiotics, chemical compounds, laser, air abrasion, access surgery, resective surgical procedures, and regenerative surgical procedures [42-49].

The results of studies on different treatment modalities for periimplantitis are inconsistent and highly controversial, and establishing a standard protocol is still impossible. Implantoplasty is a clinical procedure performed to smooth the exposed implant wires, establishing a more favourable transmucosal area. This mechanical modification of the implant surface favours the reduction of bacterial adhesion and soft tissue adaptation during the healing process

[50]. There are studies that have hypothesised that implantoplasty is effective in promoting peri-implant health and have evaluated the likelihood of successful implant survival rates after implantoplasty [9]. Mucosal recession could be a concern in the aesthetic area. However, in posterior areas or in the case of overdentures, this is not a problem and could be accepted by the patient if explained in advance [17].

Peri-implant bone resorption exposes the implant surface, treated in order to achieve better osseointegration results, as well as the implant coils. This condition represents a favourable niche for bacterial colonisation, as well as for the initiation and progression of peri-implant disease.

Mechanical modification of the implant surface in implantoplasty aims at eliminating these irregularities and achieving a smooth surface with cell viability that prevents bacterial adhesion [51]. This was highlighted in a study which demonstrated that implants treated by means of implantoplasty have less biofilm accumulation, which is also less mature than untreated implants [52].

Similarly, another study demonstrated that implantoplasty reduces the rate of bacterial growth [53]. Despite the significant reduction in roughness achieved by implantoplasty, small irregularities remain on the implant surface that promote bacterial adhesion [52,54].

Differences in implantation protocols can lead to differences in surface roughness of treated implants [55].

Several chemical forms of disinfection, including EDTA, tetracycline, hydrogen peroxide, and chlorhexidine, have been associated with implantoplasty.

However, the benefit of combining chemical disinfection with implantoplasty is still controversial [56]. Furthermore, the literature on the influence of implant surface on implant results is limited. In cases where the peri-implant defect presents a more favourable configuration for regeneration, guided surgery is combined with implantoplasty.

Clinical improvements have been described after treatment of periimplantitis with regenerative techniques associated with implantoplasty [57,58]. However, scientific evidence on the benefits of this combination is limited.

The lack of bactericidal treatments approved by accreditation agencies, due to the lengthy nature of clinical trials of different bactericidal strategies for implants [59-67], has given rise to treatments such as implantoplasty, which can be performed in any dental clinic. Bactericidal treatments under study include silver nanoparticles as an element with high bactericidal capacity [63, 66], functionalisation of titanium with organic compounds such as TESPSA and

lactoferrin, among others, and the use of titanium with organic compounds such as TESPSA and [60, 68]. The long-term behaviour of nanoparticles or the complicated and costly anchoring of organic compounds in titanium means that these treatments have not yet reached sufficient maturity for their application [69].

In the recent years, researchers from all the continents investigated, using very comprehensive microbiological assays, the bacterial aetiologies associated with dental implants. [60]

In Table 1 can be observed the most frequent bacterial species associated with periimplantitis.


**Table 1:** Bacterial species associated with periimplantitis.

First Author Year, Country	Microbial assay	Detected bacterial species	Medical importance
Zhou et al., 2022, China [57]	V4 region of the 16S rDNA Illumina HiSeq platform sequencing	<i>Actinomyces</i> <i>Streptococcus</i> <i>Fusobacterium</i> <i>Prevotella</i>	these results provide a bacteriological basis for the prevention and treatment of peri-implant mucositis
Pallos et al., 2022, Brazil [58]	V4 hypervariable region of the 16S rRNA Ion Torrent PGM System sequencing	<i>Stenotrophomonas</i> , <i>Enterococcus</i> and <i>Leuconostoc</i> genus, <i>Faecalibacterium prausnitzii</i> , <i>Haemophilus parainfluenzae</i> , <i>Prevotella copri</i> , <i>Bacteroides vulgatus</i> , <i>Bacteroides stercoris bacterial species</i>	important differences was found in participants with peri-implantitis when compared with those with healthy peri-implant sites.
Grischke et al., 2021, Germany [59]	RNAseq	16.41% of the implants were diagnosed with peri-implantitis with:  <i>Fusobacterium nucleatum subspecies animalis</i>  <i>Prevotella intermedia</i>	implants with removable dentures appear: - to be a risk indicator for peri-implantitis - to facilitate expansion of specific periodontopathogens
Ghensi et al, 2020, Italy [60]	metagenomic sequencing	peri-implantitis-related complex composed by the 7 most discriminative bacteria:  <i>P. gingivalis</i> , <i>T. forsythia</i> , <i>Treponema denticola</i> <i>P. endodontalis</i> <i>F. fastidiosum</i> species, <i>Prevotella intermedia</i> and <i>F. nucleatum</i>	the high-resolution characterization of the plaque microbiome in peri-implant diseases is the first step toward a more comprehensive understanding of the role of the microbiome in peri- implant diseases.

#### 4.2 General information on the implantoplasty technique

It is well known that one of the main problems in oral implantology is peri-implantitis, which is currently the main cause of implant unit loss. This has led to the application of the implantoplasty

procedure (Figure 1) at least in the most severe cases. However, there are no protocols, nor are there many studies on the implications of an invasive treatment such as implantoplasty [69-73].



**Figure 1.** Implantoplasty vs. physiological situation

One of the most important steps in the regenerative surgical approach to peri-implantitis is the decontamination of the implant surface.

However, implantoplasty is performed with the aim of completely removing the surface morphology of the osseointegrated implant. Thus, separate surface contamination is not a crucial issue. Despite the fact that IP is an available mechanical method for implant surface decontamination, other mechanical methods are available, and a recent randomised controlled trial showed no radiographic and clinical differences between IP and glycine air abrasion for surgical treatment of peri-implantitis [74]. Final debridement of instrumented surfaces reduces implant roughness; therefore, bacterial biofilm formation and maturation is prevented [75,76] and the risk of reinfection of treated sites is eliminated [77].

### 5.2 Hazards of invasive peri-implant therapies

“In vitro” studies have shown that performing implantoplasty on dental implants produces debris mostly in the extremely fine size range (<100 nm). Evidence from previous *in vitro* investigations has also indicated that nanosized particles may account for the largest number of debris generated; however, they represent only a very small proportion of the total volume [25,78]. These results indicate that a surgical area that visually appears “particle-free” after implantoplasty may still contain a large number of nanosized debris. It is important to note that, as previously reported in the orthopaedic literature [79, 80], smaller (submicrometric-sized) particles have been shown to be more commonly associated with macrophage activities, including increased cytokine release (e.g., TNF- $\alpha$ , IL-1 $\beta$ , IL-6, IL-8,  $\beta$ ). [81]

These extremely fine debris are more biologically harmful compared to visible coarse particles, due to biochemical mediators of inflammation, cell recruitment

and bone resorption [44]. As such, clinicians should consider alternatives to reduce the widespread release of particles generated during implantoplasty (rubber dam, high-volume evacuation, etc.). Similarly, the risks and benefits of performing implantoplasty for implant surface treatment should be weighed cautiously for each implant/patient.

Differences in number, size, mass, surface area and morphology have also been demonstrated between different manufacturers with different implant designs and surface treatments [82].

Undoubtedly, titanium processing causes a severe loss of mechanical properties due to the reduction of the dental implant surface and the significant plastic deformation produced in titanium. In addition, the effect of machining on the points of anchorage to the host tissue. From this point of view, Lozano et al consider that implantoplasty should be discouraged for cases of narrow dental implants, as there could be a risk of fracture of the dental implant [83,84].

In addition to mechanical damage, damage to chemical stability is also produced. In this contribution, we could verify the loss of corrosion resistance of the machined implant and even more so in titanium residues. As we have seen, loss in corrosion resistance is closely related to an increase in the mechanical stress accumulated in the material [36,37]. It is well known that metals with high stored energy are very sensitive to corrosion. In this study, it was found that scrap, which has the highest stored energy due to the large amount of deformation, is the most corrosive. This is of concern because these particles can remain in the surrounding tissues and can also produce titanium oxide [69].

Implantoplasty on small diameter implants produces a decrease in fracture resistance in all groups of narrow diameter titanium dental implants tested. The design of the implant-abutment connection influences the strength of fixation.

Clinicians should be aware that implantoplasty may increase the risk of fracture, especially for implants with a narrow-diameter internal connection [85].

The lower strength of narrow-diameter implants treated by means of implantoplasty should be taken into account during implant placement planning [86]. The risk of plastic deformation increases if the implant is unsplinted and there is considerable bone loss. It has been reported that a marginal bone loss of 3 mm resulted in a 37.2% reduction in mean implant strength, and the value further decreased to 53.8% for a marginal bone loss of 5 mm [87]. Therefore, implants that have undergone this type of treatment should be reviewed periodically [88].

As previously mentioned, achieving optimal surface planarity of a treated implant is the main goal of treatment. In this study, no surface analysis was performed to assess roughness. However, the milling and polishing sequence used was the same as in the study by Costa-Berenguer et al [89], where an average roughness index of  $0.1 \pm 0.02 \mu\text{m}$  was observed. This is below the established maximum value of  $0.2 \mu\text{m}$  required for the minimisation of bacterial adhesion [90]. Another limitation of this study is that we could not follow the same milling sequence for either the machined or hand-finished group. Therefore, the results for milling and polishing are not strictly comparable. Future studies including more operators and controls for milling time, pressure, and angulation of drilling during the procedure are warranted [91].

We can state with certainty, however, that many other authors have demonstrated that resective implant surgery, including implantoplasty, is an effective therapeutic approach for the treatment of peri-implantitis in terms of eliminating deep pockets, stopping inflammation and progressive marginal resorption [92-94].

However, data on crucial parameters, such as residual implant dimensions after instrumentation with

different tool kits, which might influence the resistance of implants to fracture over years of future loading, are still scarce according to the specialised literature.

There are previous studies evaluating fracture resistance which show discordant results. While Chan et al [95] found a considerable loss of fracture resistance after implantoplasty, several other publications showed no relevant loss for this parameter [96,97]. However, these latter studies did not consider thermomechanical loading, but assessed fracture loads immediately, which describes a considerable difference from the clinical situation, when after years of moderate chewing forces loading peaks could induce significant bending forces.

It should be emphasised that the results of Jorio et al. demonstrated for the first time that there was a loss of resistance for uninstrumented implants that had previously been mechanically loaded and thermally cycled compared to intact implants that had not undergone any treatment. This finding proves the unconditional importance of involving cyclic mechanical loading and thermocycling of implants before assessing fracture load.

There was no obvious difference in material loss at any measurement point for the different instrumentation techniques. This indicates that any of the instrumentation sequences evaluated could be used if only the properties of the material are considered. Damage to neighboring teeth is a clinical reality worthy of consideration in the optimisation of these clinical protocols [98].

### *5.3 Cytotoxicity and peri-implant therapy*

Toledano-Serrabona et al. [99,100] recently studied the characteristics of Ti-6Al-4V particles associated with dental implants, demonstrating a certain level of cytotoxicity and inflammatory cell reaction. Surface properties of dental implant topography [101] and the difficulty of osteoblastic growth on dental implant surfaces with

implantoplasty have also been studied. At the same time, strategies to prevent infection processes in implantoplasty treatments with the introduction of antibiotics have been studied [102]. The combination of a resective and a reconstructive surgical approach, together with locally administered antibiotic achieved a high rate of disease resolution after one year of surveillance and is a viable option for short-term management of peri-implantitis [102]. The results are encouraging but do not have long-lasting bactericidal action.

Antibiotics also have no effect on the toxicity processes of particles or on the dental implant surface in case of corrosion [60,103].

Indeed, the implantoplasty technique removes biofilms from the titanium surface to prevent periimplantitis and also preserves the osseointegrated dental implant.

However, as we have seen, it reduces corrosion resistance, increases the release of titanium ions into the environment, and reducing the cross-section of the implant through machining will lead to a reduction in mechanical properties. In this respect, it is necessary to continue research into new anodising treatment techniques, such as those published by Harraft et al [104,105] in which titanium is anodised in the form of TiO<sub>2</sub> nanotubes that can contain bactericidal drugs, or others based on citric acid [106], or with Pirahna treatments that, due to their nanotexture, prevent bacterial colonisation [107].

## **6. Microbiological perspectives in peri-implant disease research**

Lozano et al pointed out that one aspect that requires further study is how the titanium surface will behave towards bacteria after implantation. In principle, everything suggests that having a surface with a lower roughness than that of the implant as received will make biofilm formation more difficult [62,63]. It is well known that the roughness of the dental

implant, which is fundamental for osseointegration and fixation of the implant onto the bone, is unfavourable for bacterial colonisation [108,109]. Therefore, the reduced roughness caused by the implant machining process would favour the reduction of plaque adhesion. Similarly, the lack of a certain roughness should be studied in terms of the possibility of osteoblasts to adhere, proliferate and differentiate on this surface. This will certainly be a disadvantage for new bone formation.

For all these reasons, Lozano et al suggest that it is imperative to continue studying these biological aspects on titanium surfaces that have undergone implantoplasty [69].

Cytotoxicity and peri-implant therapy are an important issue associated with implantoplasty.

## 7. Conclusions

Implantoplasty may be considered by many specialists as modern alternative of treatment for peri-implant pathological processes. Within the limitations of the current paper, we have emphasised the key role of some elements of etiopathogenicity such as dental plaque and smoking. In the same time we pointed out that molecular biology mechanisms involved in peri-implant pathology still need to be explored and we detailed possible particular hazards of invasiveness certain peri-implant therapies.

Using sensitive and specific methods (e.g., sequencing), numerous bacterial species associated with dental implants can be successfully identified at early stages and thus, reduce the failure rate of the implants. Therefore, such techniques like implantoplasty require additional studies and careful prospective evaluation especially from the microbiological perspective, in order to significantly increase the survival rate of dental implants.

## References

1. Renvert S, Persson GR, Piri FQ, Camargo PM: Peri-implant health, peri-implant mucositis, and peri-implantitis: Case definitions and diagnostic considerations. *J Periodontol* 89:S304-S312, 2018
2. Lee CT, Huang YW, Zhu L, Weltman R: Prevalences of peri-implantitis and peri-implant mucositis: systematic review and meta-analysis. *J Dent* 62:1-12, 2017
3. Schwarz F, Derks J, Monje A, Peri-implantitis Wang HL: *J Periodontol* 89:S267-SS90, 2018
4. Popa, CG; Luchian, I; (...); Martu, S. ELISA Evaluation of RANKL Levels in Gingival Fluid in Patients with Periodontitis and Occlusal Trauma. *Rev. de Chimie*. 2018, 69 (6):1578-1580.
5. Luchian I, Goriuc A, Sandu D, Covasa M. The Role of Matrix Metalloproteinases (MMP-8, MMP-9, MMP-13) in Periodontal and Peri-Implant Pathological Processes. *Int J Mol Sci*. 2022 Feb 4;23(3):1806.
6. Luchian I, Martu I, Ioanid N, Goriuc A, Vata I, Martu-Stefanache A, Hurjui L, Tatarciuc M, Matei MN, Martu S. Salivary IL1 $\beta$ : a Biochemical Marker that Predicts Periodontal Disease in Orthodontic Treatment. *Rev Chim (Bucharest)*, 2016; 67(12):2479-83.
7. Luchian I, Martu I, Goriuc A, Vata I, Hurjui L, Matei MN, Martu S. Salivary PGE2 as a Potential Biochemical Marker during Orthodontic Treatment Associated with Periodontal Disease. *Rev Chim (Bucharest)*, 2016; 67(10):2119-23.
8. Luchian AI, Martu I, Goriuc A, Martu C, Beldiman A, Martu S. Changes in biochemical parameters associated with periodontal disease. *Rev. Chim. (Bucharest)*, 2016, 67 (6):1073-1075
9. Esteves Lima RP, Abreu LG, Belém FV, Pereira GHM, Brant RA, Costa FO: Is Implantoplasty Efficacious at Treating Peri-Implantitis? A Systematic Review and Meta-Analysis. *J Oral Maxillofac Surg*. 2021, 79(11):2270-2279.
10. Ferreira, S.D.; Silva, G.L.; Cortelli, J.R.; Costa, J.E.; Costa, F.O: Prevalence and risk variables for peri-implant disease in Brazilian subjects. *J. Clin. Periodontol*. 2006, 33: 929-935,

11. Solomon SM, Chiscop I, Moisei M, Badescu AC, Jelihovschi I, Martu-Stefanache A, Teusan A, **Martu Silvia**, Iancu LS. Evaluation of DNA extraction. methods from saliva as a source of PCR - amplifiable genomic DNA. *Rev. Chim. (Bucharest)*, 2015,67 (12):2101-2103
12. Costa, F.O.; Ferreira, S.D.; Cortelli, J.R.; Lima, R.P.E.; Cortelli, S.C.; Cota, L.O.M. Microbiological profile associated with peri-implant diseases in individuals with and without preventive maintenance therapy: A 5-year follow-up. *Clin. Oral Investig.* 2019, 23:3161–3171,
13. Ursu RG, Iancu LS, Porumb-Andrese E, Damian C, Cobzaru RG, Nichitean G, Ripa C, Sandu D, Luchian I.: Host MRNA analysis of periodontal disease patients positive for porphyromonas gingivalis, aggregatibacter actinomycetemcomitans and tannerella forsythia, *Int J Mol Sci*, 2022, 23(17).
14. Dalago, H.R.; Filho, G.S.; Rodrigues, M.A.; Renvert, S.; Bianchini, M.A.: Risk indicators for Peri-implantitis. A cross-sectional study with 916 implants. *Clin. Oral Implants Res.* 2017, 28:144–150.
15. Martu MA, Solomon SM, Sufaru IG, Jelihovschi I, Martu S, Rezus E, Surdu AE, Onea RM, Grecu GP, Foia L. Study on the Prevalence of Periodontopathogenic Bacteria in Serum and Subgingival Bacterial Plaque in Patients with Rheumatoid Arthritis. *Rev. Chim. (Bucharest)*. 2017; 68(8):1946-1949
16. Mocanu, RC; Martu, MA; (...); Tatarciuc, M. Microbiologic Profiles of Patients with Dental Prosthetic Treatment and Periodontitis before and after Photoactivation Therapy-Randomized Clinical Trial, *Microorganisms*, Apr 2021, 9 (4).
17. Martins O, Sahrman P, Ramos J, Caramelo F, Matos S, Baptista IP. Implantoplasty Improves Clinical Parameters over a 2-Year Follow-Up: A Case Series. *Medicina (Kaunas)* 2022, 58(1):113,
18. Jung RE, Zembic A, Pjetursson BE, Zwahlen M, Thoma DS: Systematic review of the survival rate and the incidence of biological, technical, and aesthetic complications of single crowns on implants reported in longitudinal studies with a mean follow-up of 5 years. *Clin Oral Implants Res*, 2012, 23:2-21,
19. Luca AS, Luca MC, Ursu RG, Iancu LS. The Study Of VHB Genotypes In A Population - A Pilot Study, *Rev Chim*, 2016, 67(3):500-502.
20. Romeo E, Ghisolfi M, Carmagnola D: Peri-implant diseases. A systematic review of the literature. *Minerva Stomatol*, 2004, 53:215- 230,
21. Heitz-Mayfield, L.J. Peri-implant diseases: Diagnosis and risk indicators. *J. Clin. Periodontol.* 2008, 35 (Suppl. 8):292–304,
22. Aglietta, M.; Siciliano, V.I.; Rasperini, G.; Cafiero, C.; Lang, N.P.; Salvi, G.E. A 10-year retrospective analysis of marginal bone-level changes around implants in periodontally healthy and periodontally compromised tobacco smokers. *Clin. Oral Implants Res.* 2011, 22:47–53,
23. Englezos, E.; Cosyn, J.; Koole, S.; Jacquet, W.; De Bruyn, H. Resective Treatment of Peri-implantitis: Clinical and Radiographic Outcomes After 2 Years. *Int. J. Periodontics Restor. Dent.* 2018, 38:729–735,
24. Magone, K.; Luckenbill, D.; Goswami, T. Metal ions as inflammatory initiators of osteolysis. *Arch. Orthop. Trauma Surg.* 2015, 135:683–695,
25. Shanbhag, A.S.; Jacobs, J.J.; Glant, T.T.; Gilbert, J.L.; Black, J.; Galante, J.O. Composition and morphology of wear debris in failed uncemented total hip replacement. *J. Bone Jt. Surg. Br.* 1994,76:60–67,
26. Martu I, Luchian I, Goriuc A, Tatarciuc M, Ioanid N, Cioloca DP, Bodnariu GE, Martu C. Comparative Analysis of some Antioxidant Markers in Periodontal Disease. *Rev Chim (Bucharest)*, 2016; 67(7):1378-81.
27. Teodorescu AC, Martu I, Teslaru S, Kappenberg-Nitescu DC, Goriuc A, Luchian I, Martu MA, Solomon SM, Mârțu S. Assessment of Salivary Levels of RANKL and OPG in Aggressive versus Chronic Periodontitis. *J of Immunology Res*, Hindawi, 2019, Article ID 6195258, 6 pg. <https://doi.org/10.1155/2019/6195258>
28. Archibeck, M.J.; Jacobs, J.J.; Roebuck, K.A.; Glant, T.T. The basic science of periprosthetic osteolysis. *Instr. Course. Lect.* 2001,50:185–195,
29. Franchi, M.; Bacchelli, B.; Martini, D.; de Pasquale, V.; Orsini, E.; Ottani, V.; Fini, M.; Giavaresi, G.; Giardino, R.; Ruggeri, A. Early detachment of titanium particles from various different surfaces of endosseous dental implants. *Biomaterials* 25, 2004:2239–2246,

30. Schliephake, H.; Reiss, G.; Urban, R.; Neukam, F.W.; Guckel, S. Metal release from titanium fixtures during placement in the mandible: An experimental study. *Int. J. Oral Maxillofac. Implant.* 1993, 8:502–511,
31. Senna, P.; Antoninha Del Bel Cury, A.; Kates, S.; Meirelles, L. Surface Damage on Dental Implants with Release of Loose Particles after Insertion into Bone. *Clin. Implant. Dent. Relat. Res.* 2015, 17:681–692,
32. Wilson, T.G., Jr.; Valderrama, P.; Burbano, M.; Blansett, J.; Levine, R.; Kessler, H.; Rodrigues, D.C. Foreign Bodies Associated With Peri-Implantitis Human Biopsies. *J. Periodontol.* 2015, 86:9–15,
33. Flatebø, R.S.; Johannessen, A.C.; Grønningsæter, A.G.; Bøe, O.E.; Gjerdet, N.R.; Grung, B.; Leknes, K.N. Host Response to Titanium Dental Implant Placement Evaluated in a Human Oral Model. *J. Periodontol.* 2006,77:1201–1210,
34. Olmedo, D.G.; Paparella, M.L.; Spielberg, M.; Brandizzi, D.; Guglielmotti, M.B.; Cabrini, R.L. Oral mucosa tissue response to titanium cover screws. *J. Periodontol.* 2012, 83:973–980,
35. Ingham, E.; Fisher, J. The role of macrophages in osteolysis of total joint replacement. *Biomaterials*, 2005,26:1271–1286,
36. Ursu RG, Luca CM, Luca AS, Toader E, Simion L, Iancu LS.: Laboratory diagnosis for optimize therapy of B hepatitis virus infection by using biochemical and molecular biology methods, *Rev Chim*, 2016, 67(12):2614-2617,
37. Nemescu D, Ursu RG<sup>1</sup>, Nemescu ER, Negura L.: Heterogeneous distribution of fetal microchimerism in local breast cancer environment, *Plos One*, 2016, 11(1): art. no e0147675,
38. Porumb-Andrese E, Scutariu MM, Luchian I, Schreiner TG, Martu I, Porumb V, Popa CG, Sandu D, Ursu RG.: Molecular profile of skin cancer, *Appl Sci-Basel*, 2021, 11(19): art. no 9142,
39. Purdue, P.E.; Koulouvaris, P.; Nestor, B.J.; Sculco, T.P. The central role of wear debris in periprosthetic osteolysis. *HSS J. Hosp. Spec. Surg.* 2006, 2:102–113,
40. Wimhurst, J.A.; Brooks, R.A.; Rushton, N. Inflammatory responses of human primary macrophages to particulate bone cements in vitro. *J. Bone Jt. Surg. Br.* 2001, 83:278–282,
41. Catelas, I.; Huk, O.L.; Petit, A.; Zukor, D.J.; Marchand, R.; Yahia, L. Flow cytometric analysis of macrophage response to ceramic and polyethylene particles: Effects of size, concentration, and composition. *J. Biomed. Mater. Res.* 1998, 41:600–607,
42. Anitua, E.A. Enhancement of osseointegration by generating a dynamic implant surface. *J. Oral Implantol.* 2006,32:72–76,
43. De Maeztu, M.A.; Alava, J.I.; Gay-Escoda, C. Ion implantation: Surface treatment for improving the bone integration of titanium and Ti6Al4V dental implants. *Clin. Oral Implant. Res.* 2003,14:57–62,
44. Berglundh, T.; Gotfredsen, K.; Zitzmann, N.U.; Lang, N.P.; Lindhe, J. Spontaneous progression of ligature induced peri-implantitis at implants with different surface roughness: An experimental study in dogs. *Clin. Oral Implant. Res.* 2007,18:655–661,
45. Ramanauskaite A, Daugela P, Faria de Almeida R, Saulacic N: Surgical non-regenerative treatments for peri-implantitis: a systematic review. *J Oral Maxillofac Res*, 2016, 7(3):e14,
46. Ursu RG, Luchian I, Ghetu, N, Costan VV, Stamatin O, Palade OD, Damian C, Iancu LS, Porumb-Andrese E.: Emerging oncogenic viruses in head and neck cancers from romanian patients, *Appl Sci-Basel*, 2021,11(19): art. no 9356,
47. Becker A, Zogakis I, Luchian I, Chaushu S. Surgical exposure of impacted canines: Open or closed surgery?. *Semin Orthod*, 2016; 22(1):27-33.
48. Martu, M.-A.; Surlin, P.; Lazar, L.; Maftai, G.A.; Luchian, I.; Gheorghe, D.-N.; Rezus, E.; Toma, V.; Foia, L.-G. Evaluation of Oxidative Stress before and after Using Laser and Photoactivation Therapy as Adjuvant of Non-Surgical Periodontal Treatment in Patients with Rheumatoid Arthritis. *Antioxidants* 2021,10, 226.
49. Martu, M.A., Maftai, G.A., Luchian, I., Popa, C., Filioreanu, A.M., Tatarciuc, D. Foia, L.G. Wound healing of periodontal and oral tissues: Part II—Patho-physiological conditions and metabolic diseases. *Rom. J. Oral Rehabil*, 2020, 12(4):30-40

50. Bianchini MA, Galarraga-Vinueza ME, Apaza-Bedoya K, De Souza JM, Magini R, Schwarz F: Two to 6-year disease resolution and marginal bone stability rates of a modified resective-implantoplasty therapy in 32 peri-implantitis cases. *Clin Implant Dent Relat Res*, 2019, 21:758-765,
51. Schwarz F, John G, Becker J: The influence of implantoplasty on the diameter, chemical surface composition, and biocompatibility of titanium implants. *Clin Oral Investig*. 2016, 21:1-7,
52. Azzola F, Ionescu AC, Ottobelli M, et al. Biofilm formation on dental implant surface treated by implantoplasty: an in situ study. *Dent J*, 2020, 8:40,
53. Geremias TC, Montero JFD, Magini RS, Schuldt Filho G, Magalhães Jr EB, Bianchini MA: Biofilm analysis of retrieved dental implants after different peri-implantitis treatments. *Case Rep Dent* 2017:1-5,
54. Costa-Berenguer X, Garcia-Garcia M, Sanchez-Torres A, Sanz-Alonso M, Figueiredo R, Valmaseda-Castellon E: Effect of implantoplasty on fracture resistance and surface roughness of standard diameter dental implants. *Clin Oral Impl Res*, 2018, 29:46-54,
55. Ramel CF, Lussi A, Ozcan M, Jung RE, Hammerle CHF, Thoma DS: Surface roughness of dental implants and treatment time using six different implantoplasty procedures. *Clin Oral Impl Res* 2016, 27:776-781,
56. Valderrama P, Wilson T G Jr.: Detoxification of implant surfaces affected by peri-implant disease: an overview of surgical methods. *Int J Dent* 2013:740680,
57. Matarasso S, Iorio Siciliano V, Aglietta M, Andreuccetti G, Salvi GE: Clinical and radiographic outcomes of a combined resective and regenerative approach in the treatment of peri-implantitis: a prospective case series. *Clin Oral Implants Res*, 2014, 25:761-767,
58. Schwarz F, John G, Schmucker A, Sahm N, Becker J: Combined surgical therapy of advanced peri-implantitis evaluating two methods of surface decontamination: a 7-year follow-up observation. *J Clin Periodontol*, 2017,44:337-342,
59. Nicolae V, Chiscop I, Ibric Cioranu VS, Martu MA, Luchian AI, Martu S, Solomon SM. The effects of the therapy with LED photo-activated toluidine blue o in patients with peri-implantitis.. *Rev. Chim. (Bucharest)*, 2015,67 (12):2121-2123
60. Luchian, I; Goriuc, A; (...); Covasa, M. Clindamycin as an Alternative Option in Optimizing Periodontal Therapy, *Antibiotics-Basel*, 2021,10(7):
61. Ursu RG, Onofriescu M, Luca A, Prisecariu LJ, Salceanu SO, Nemescu D, Iancu LS.: The need for cervical cancer control in hiv-positive and hiv-negative women from romania by primary prevention and by early detection using clinically validated HPV/DNA tests, *Plos One*, 2015,10(7): art. no e0132271,
62. Godoy-Gallardo, M.; Manzanares-Céspedes, M.C.; Sevilla, P.; Nart, J.; Manzanares, N.; Manero, J.M.; Gil, F.J.; Boyd, S.K.; Rodríguez, D. Evaluation of bone loss in antibacterial coated dental implants: An experimental study in dogs. *Mater. Sci. Eng. C Mater. Biol. Appl.* , 2016, 69:538–545
63. Hoyos, M.; Velasco, F.; Ginebra, M.P.; Manero, J.M.; Gil, F.J.; Mas-Moruno, C. Regenerating bone via multifunctional coatings: The blending of cell integration and bacterial inhibition properties on the Surface of biomaterials. *ACS Appl. Mater. Interf.* 2019, 11:36449–36457,
64. Garcia-Godoy, M.; Rodríguez-Hernández, A.; Delgado, L.; Manero, J.M.; Gil, F.J.; Rodriguez, D. Silver deposition on titanium surface by electrochemical anodizing process reduces bacterial adhesion of *Streptococcus sanguinis* and *Lactobacillus salivarius*. *Clin.Oral Impl. Res.* 2014, 26:1170–1179,
65. Punset, M.; Villarrasa, J.; Nart, J.; Manero, J.M.; Bosch, B.; Padrós, R.; Perez, R.A.; Gil, F.J. Citric Acid Passivation of Titanium Dental Implants for Minimizing Bacterial Colonization Impact. *Coatings*, 2021,11, 214,
66. Godoy-Gallardo, M.; Guillem-Marti, J.; Sevilla, P.; Manero, J.M.; Gil, F.J.; Rodriguez, D. Anhydride-functional silane immobilized onto titanium surfaces induces osteoblast cell differentiation and reduces bacterial adhesion and biofilm formation. *Mater. Sci. Eng. C Mater. Biol. Appl.* 2016, 59:524–532,
67. Zhao, L.; Chu, P.K.; Zhang, Y.; Wu, Z. Antibacterial coatings on titanium implants. *J. Biomed. Mater. Res. B Appl. Biomater.* 2009, 91:470–480,

68. Buxadera-Palomero, J.; Godoy-Gallardo, M.; Molmeneu, M.; Punset, M.; Gil, F.J. Antibacterial properties of triethoxysilylpropyl succinic anhydride silane (TESPSA) on Titanium dental implants. *Polymers* 2020, 12, 773,
69. Lozano, P., Peña, M., Herrero-Climent, M., Rios-Santos, J. V., Rios-Carrasco, B., Brizuela, A., & Gil, J.: Corrosion Behavior of Titanium Dental Implants with Implantoplasty. *Materials (Basel, Switzerland)*, 2022, 15(4):1563,
70. Zhou N, Huang H, Liu H, Li Q, Yang G, Zhang Y, Ding M, Dong H, Mou Y.: Microbiota analysis of peri-implant mucositis in patients with periodontitis history. *Clin Oral Investig.* 2022, Oct;26(10):6223-6233,
71. Pallos D, Sousa V, Feres M, Retamal-Valdes B, Chen T, Curtis M, Boaventura RM, Tanaka MH, Salomão GVDS, Zanella L, Tozetto-Mendoza TR, Schwab G, Franco LAM, Sabino EC, Braz-Silva PH, Shibli JA.: Salivary Microbial Dysbiosis Is Associated With Peri-Implantitis: A Case-Control Study in a Brazilian Population. *Front Cell Infect Microbiol.* 2022 Jan, 5;11:696432,
72. Grischke J, Szafranski SP, Muthukumarasamy U, Haeussler S, Stiesch M.: Removable denture is a risk indicator for peri-implantitis and facilitates expansion of specific periodontopathogens: a cross-sectional study. *BMC Oral Health.* 2021, Apr 1;21(1):173,
73. Ghensi P, Manghi P, Zolfo M, Armanini F, Pasolli E, Bolzan M, Bertelle A, Dell'Acqua F, Dellasega E, Waldner R, Tessarolo F, Tomasi C, Segata N.: Strong oral plaque microbiome signatures for dental implant diseases identified by strain-resolution metagenomics. *NPJ Biofilms Microbiomes.* 2020, Oct 30;6(1):47,
74. Lasserre, J.F.; Brex, M.C.; Toma, S. Implantoplasty Versus Glycine Air Abrasion for the Surgical Treatment of Peri-implantitis: A Randomized Clinical Trial. *Int. J. Oral Maxillofac. Implants*, 2020, 35:197–206,
75. Subramani, K.; Jung, R.E.; Molenberg, A.; Hammerle, C.H. Biofilm on dental implants: A review of the literature. *Int. J. Oral Maxillofac. Implants*, 2009, 24:616–626,
76. Quirynen, M.; van der Mei, H.C.; Bollen, C.M.; Schotte, A.; Marechal, M.; Doornbusch, G.I.; Naert, I.; Busscher, H.J.; van Steenberghe, D. An in vivo study of the influence of the surface roughness of implants on the microbiology of supra- and subgingival plaque. *J. Dent. Res.*, 1993, 72:1304–1309,
77. Schwarz, F.; Sahm, N.; Iglhaut, G.; Becker, J. Impact of the method of surface debridement and decontamination on the clinical outcome following combined surgical therapy of peri-implantitis: A randomized controlled clinical study. *J. Clin. Periodontol*, 2011, 38:276–284,
78. Ingram, J.H.; Stone, M.; Fisher, J.; Ingham, E. The influence of molecular weight, crosslinking and counterface roughness on TNF-alpha production by macrophages in response to ultra high molecular weight polyethylene particles. *Biomaterials*, 2004, 25: 3511–3522,
79. Al-Saffar, N.; Khwaja, H.A.; Kadoya, Y.; Revell, P.A. Assessment of the Role of GM-CSF in the Cellular Transformation and the Development of Erosive Lesions around Orthopaedic Implants. *Am. J. Clin. Pathol.* 1996, 105:628–639,
80. Sabokbar, A.; Rushton, N. Role of inflammatory mediators and adhesion molecules in the pathogenesis of aseptic loosening in total hip arthroplasties. *J. Arthroplast.*, 1995,10:810–816,
81. Nicolae, V; Neamtu, B; (...); Cioranu, VSI, The Comparative Evaluation of Salivary Biomarkers (Calcium, Phosphate, Salivary pH) in Cario-resistance Versus Cario-activity. *Rev. de Chimie*, 2016, 67 (4):821-824
82. Wu, X., Cai, C., Gil, J., Jantz, E., Al Sakka, Y., Padiá-Molina, M., & Suárez-López Del Amo, F.: Characteristics of Particles and Debris Released after Implantoplasty: A Comparative Study. *Materials (Basel, Switzerland)*, 2022, 15(2):602,
83. Velasco-Ortega, E.; Flichy-Fernández, A.; Punset, M.; Jiménez-Guerra, A.; Manero, J.M.; Gil, F.J. Fracture and Fatigue of Titanium Narrow Dental Implants: New Trends in Order to Improve the Mechanical Response. *Materials*, 2019,12:3728,
84. Pérez, R.A.; Gargallo, J.; Altuna, P.; Herrero-Climent, M.; Gil, F.J. Fatigue of Narrow Dental Implants: Influence of the Hardening Method. *Materials*, 2020,13:1429,
85. Camps-Font, O., González-Barnadas, A., Mir-Mari, J., Figueiredo, R., Gay-Escoda, C., & Valmaseda-Castellón, E.: Fracture resistance after implantoplasty in three implant-abutment connection designs. *Medicina oral, patologia oral y cirugía bucal*, 2020, 25(5): e691–e699,

86. Suarez F, Monje A, Galindo-Moreno P, Wang HL. Implant surface detoxification: a comprehensive review. *Implant Dent.*; 2013, 22(5):465–73,
87. Gherke SA, Souza Dos Santos Vianna M, Dedavid BA. Influence of bone insertion level of the implant on the fracture strength of different connection designs: an in vitro study. *Clin Oral Invest.*; 2014, 18(3):715–20,
88. Chan HL, Oh WS, Ong HS, Fu JH, Steigmann M, Sierraalta M, et al. Impact of implantoplasty on strength of the implant-abutment complex. *Int J Oral Maxillofac Implants.*; 2013, 28(6):1530–5,
89. Costa-Berenguer X, Garcia-Garcia M, Sanchez-Torres A, Sanz-Alonso M, Figueiredo R, Valmaseda-Castellon E. Effect of implantoplasty on fracture resistance and surface roughness of standard diameter dental implants. *Clin Oral Implants Res.*; 2018, 29(1):46–54,
90. Bollen CM, Papaioanno W, Van Eldere J, Schepers E, Quirynten M, Van Steenberghe D. The influence of abutment surface roughness on plaque accumulation and peri-implant mucositis. *Clin Oral Implants Res.*; 1996, 7(3): 201–11,
91. Diéguez-Pereira, M., Chávarri-Prado, D., Viteri-Agustín, I., Montalban-Vadillo, O., Pérez-Pevida, E., & Brizuela-Velasco, A.: . Effect of implantoplasty on the elastic limit of dental implants of different diameters. *International journal of implant dentistry*, 2021, 7(1), 88,
92. Romeo, E., Ghisolfi, M., Murgolo, N., Chiapasco, M., Lops, D., Vogel, G.: Therapy of peri-implantitis with resective surgery. A 3-year clinical trial on rough screw-shaped oral implants. Part I: clinical outcome. *Clin Oral Implants Res.* 2005,16, 1:9-18,
93. Romeo, E., Lops, D., Chiapasco, M., Ghisolfi, M., Vogel, G.:Therapy of peri-implantitis with resective surgery. A 3-year clinical trial on rough screw-shaped oral implants. Part II: radiographic outcome. *Clin Oral Implants Res.* 2007, 18, 2:179-187,
94. Berglundh, T., Wennström, J.L., Lindhe, J.: Long-term outcome of surgical treatment of peri-implantitis. A 2-11-year retrospective study. *Clin Oral Implants Res.* 2018, 29, 4:404-410,
95. Chan H.L., Oh, W.S., Ong, H.S., Fu, J.H., Steigmann, M., Sierraalta, M., Wang, H.L.: Impact of implantoplasty on strength of the implant-abutment complex. *Int J Oral Maxillofac Implants.* 2013, 28, 6:1530-1535,
96. Sahrman, P., Luso, S., Mueller, C., Ender, A., Attin, T., Stawarczyk, B., Schmidlin, P.R.: Titanium Implant Characteristics After Implantoplasty: An In Vitro Study on Two Different Kinds of Instrumentation. *Int J Oral Maxillofac Implants.* 2019, 34, 6:1299-1305,
97. Costa-Berenguer, X., García-García, M., Sánchez-Torres, A., Sanz-Alonso, M., Figueiredo, R., Valmaseda-Castellón, E.: Effect of implantoplasty on fracture resistance and surface roughness of standard diameter dental implants. *Clin Oral Implants Res.* 2018, 29, 1:46-54,
98. Jorio, I. C., Stawarczyk, B., Attin, T., Schmidlin, P. R., & Sahrman, P.: Reduced fracture load of dental implants after implantoplasty with different instrumentation sequences. An in vitro study. *Clinical oral implants research* 2021, 2(8):881–892,
99. Toledano-Serrabona, J.; Gil, F.J.; Camps-Font, O.; Valmaseda-Castellon, E.; Gay-Escoda, C.; Sánchez-Garcés, M.A. Physicochemical and Biological Characterization of Ti6Al4V particles obtained by Implantoplasty: An In vivo study. Part I. *Materials*, 2021,14:6507,
100. Toledano-Serrabona, J.; Sánchez-Garcés, M.A.; Gay-Escoda, C.; Valmaseda-Castellon, E.; Camps-Font, O.; Verdeguer, P.; Molmeneu, M.; Gil, F.J. Mechanical properties and corrosión behavior of Ti6Al4V particles obtained by Implantoplasty. An in vivo study. Part II. *Materials*, 2021,14:6519,
101. Burgueño-Barris, G.; Camps-Font, O.; Figueiredo, R.; Valmaseda-Castellón, E. The Influence of Implantoplasty on Surface Roughness, Biofilm Formation, and Biocompatibility of Titanium Implants: A Systematic Review. *J. Oral Maxil. Impl.*, 2021, 36, 111–119,
102. Ursu RG, Luca CM, Luca AS, Toader E, Simion L, Iancu LS.: Laboratory diagnosis for optimize therapy of b hepatitis virus infection by using biochemical and molecular biology methods, *Rev Chim*, 2016, 67(12): 2614-2617,
103. Spiridon I, Darie Nita RN, Kozlowski M, Nechita A, Ursu RG.: Influence of accelerated weathering on the performance of polylactic acid based materials, *Cellulose Chem Technol*, 2016, 50(5-6): 629-635,
104. González Regueiro, I.; Martínez Rodríguez, N.; Barona Dorado, C.; Sanz-Sánchez, I.; Montero, E.; Ata-Ali, J.; Duarte, F.; Martínez- González, J.M. Surgical approach combining implantoplasty

- and reconstructive therapy with locally delivered antibiotic in the treatment of peri-implantitis: A prospective clinical case series. *Clin. Implant. Dent. Relat. Res.*, 2021, 23:864–873,
105. Sarraf, M.; Sukiman, N.L.; Bushroa, A.R.; Nasiri-Tabrizi, B.; Dabbagh, A.; Abu Kasim, N.H.; Basirun, W.J. In vitro bioactivity and corrosion resistance enhancement of Ti-6Al-4V by highly ordered TiO<sub>2</sub> nanotube arrays. *J. Aust. Cer. Soc.*, 2019, 55:187–200,
106. Verdeguer, P.; Gil, J.; Punset, M.; Manero, J.M.; Nart, J.; Vilarrasa, J.; Ruperez, E. Citric Acid in the Passivation of Titanium Dental Implants: Corrosion Resistance and Bactericide Behavior. *Materials*, 2022, 15:545,
107. Cruz, N.; Gil, J.; Punset, M.; Manero, J.M.; Tondela, J.P.; Verdeguer, P.; Aparicio, C.; R perez, E. Relevant Aspects of Piranha Passivation in Ti6Al4V Alloy Dental Meshes. *Coatings*, 2022,12:154,
108. Aparicio, C.; Rodriguez, D.; Gil, F.J. Variation of roughness and adhesion strength of deposited apatite layers on titanium dental implants. *Mater. Sci. Eng. C*, 2011, 31:320–324,
109. Von Wilmowsky, C.; Moest, T.; Nkenke, E.; Stelzle, F.; Schlegel, K.A. Implants in bone: Part I. A current overview about tissue response, surface modifications and future perspectives. *Oral Maxillofac. Surg.*, 2013,18:243–257,